

---

---

**Sterilization of medical devices —  
Information to be provided by the  
manufacturer for the processing of  
resterilizable medical devices**

*Stérilisation des dispositifs médicaux — Informations devant être  
fournies par le fabricant pour le processus de restérilisation des  
dispositifs médicaux*



**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 17664 was prepared by the European Committee for Standardization (CEN) in collaboration with Technical Committee ISO/TC 198, *Sterilization of health care products*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

Throughout the text of this document, read “...this European Standard...” to mean “...this International Standard...”.

For the purposes of this International Standard, the CEN annex regarding fulfilment of European Council Directives has been removed.

<b>Contents</b>	<b>Page</b>
Foreword .....	v
Introduction.....	vi
1 Scope .....	1
2 Terms and definitions .....	1
3 Information to be provided by the medical device manufacturer .....	3
3.1 Reprocessing instructions .....	3
3.2 Limitations and restrictions on reprocessing .....	3
3.3 Preparation at the point of use prior to processing .....	3
3.4 Preparation before cleaning.....	4
3.5 Cleaning .....	4
3.6 Disinfection .....	4
3.7 Drying .....	5
3.8 Inspection, maintenance and testing .....	5
3.9 Packaging.....	6
3.10 Sterilization .....	6
3.11 Storage .....	6
4 Presentation of the information .....	6
5 Validation of the reprocessing information provided.....	7
6 Risk analysis.....	7
Annex A (informative) Commonly utilized reprocessing methods.....	8
A.1 General .....	8
A.2 Matrix for identifying methods of processing .....	9
Annex B (informative) An example of reprocessing instructions for reusable medical devices .....	13
Bibliography.....	18

## Foreword

This document (EN ISO 17644:2004) has been prepared by Technical Committee CEN/TC 204 "Sterilization of medical devices", the secretariat of which is held by BSI, in collaboration with Technical Committee ISO/TC 198 "Sterilization of health care products".

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2004, and conflicting national standards shall be withdrawn at the latest by September 2004.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

Informative annexes A and B are attached to this document.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Luxembourg, Malta, Netherlands, Norway, Portugal, Slovakia, Spain, Sweden, Switzerland and the United Kingdom.

## Introduction

This standard applies to those medical devices which are intended for multiple use and require processing to take them from their state at the end of one use to the state of being sterile and ready for their subsequent use. Some medical devices supplied non-sterile but intended to be used in a sterile state, will also require similar treatment.

## 1 Scope

This standard specifies the information to be provided by the medical device manufacturer on the processing of medical devices claimed to be re-sterilizable and medical devices intended to be sterilized by the processor.

This standard specifies requirements for the information to be provided by the medical device manufacturer, so that the medical device can be processed safely and will continue to meet its performance specification.

Requirements are specified for processing that consists of all or some of the following activities:

- preparation at the point of use;
- preparation, cleaning, disinfection;
- drying;
- inspection, maintenance and testing;
- packaging;
- sterilization;
- storage.

When providing instructions for these activities, medical device manufacturers are expected to be aware of the training and knowledge of procedures, and of the processing equipment available to the persons likely to be responsible for processing. It is likely that some processing procedures will be generic and well known and will use equipment and consumables conforming to recognized standards. In this case, a reference in the instructions is all that is required. For those medical devices where instructions for use are not required to accompany the medical device, other means of communicating the information can be used, e.g. user manuals, symbols or wall charts supplied separately.

This standard excludes textile devices used in patient draping systems or surgical clothing.

**NOTE** The principles of this standard may be applied when considering the information to be supplied with medical devices which only require disinfection prior to re-use.

## 2 Terms and definitions

For the purposes of this European Standard, the following terms and definitions apply.

### 2.1

#### **chemical**

formulation of compounds intended for use in reprocessing

**NOTE** This includes, for example, detergents, surfactants, rinse aids, disinfectants, enzymatic cleaners, sterilants.

### 2.2

#### **cleaning**

removal of contamination from an item to the extent necessary for further processing or for intended use

## 2.3

### **disinfection**

process used to reduce the number of viable microorganisms on a product to a level previously specified as appropriate for its further handling or use

## 2.4

### **manual cleaning**

cleaning without the use of a washer-disinfector

## 2.5

### **manufacturer**

organization with responsibility for the design, manufacture, packaging and labelling of a device before it is placed on the market under its own name, regardless of whether these operations are carried out by that person himself or on its behalf by a third party

## 2.6

### **processing**

activity including cleaning, disinfection and sterilization, necessary to prepare a new or used medical device for its intended use

## 2.7

### **processor**

organization with the responsibility for carrying out the actions necessary to prepare a new or used device for its intended use

## 2.8

### **sterilant**

chemical which has properties to destroy micro-organisms including viruses, when used at correct dilution/dose and applied for recommended exposure time

## 2.9

### **sterile**

free from all viable microorganisms

## 2.10

### **sterilization**

process used to render a device free from all forms of viable microorganisms

NOTE In a sterilization process, the nature of microbiological death is described by an exponential function. Therefore, the presence of microorganisms on any individual item may be expressed in terms of probability. Whilst this probability may be reduced to a very low number, it can never be reduced to zero. (See ISO 11134). This probability can only be assured for validated processes.

## 2.11

### **validation**

documented procedure for obtaining, recording and interpreting the results required to establish that a process will consistently yield product complying with predetermined specifications

## 2.12

### **verification**

confirmation by examination and provision of objective evidence that specified requirements have been fulfilled



**2.13****washer-disinfectors**

machine intended to clean and disinfect medical devices and other articles used in the context of medical, dental, pharmaceutical and veterinary practice

NOTE 1 This type of machine does not include those designed specifically to wash linen or clothing. Machines intended to sterilize, or designated as sterilizers, are specified in other standards e.g. EN 285.

NOTE 2 Preliminary standards for washer-disinfectors (prEN 15883) are being prepared in an ISO-CEN project.

**3 Information to be provided by the medical device manufacturer****3.1 Reprocessing instructions**

At least one validated method for reprocessing the medical device shall be specified.

The following information shall be stated where it is critical to the maintenance of the intended function of the medical device and the safety of the user(s) and the patient:

- details of process steps;
- a description of special equipment and/or accessories;
- specification of process parameters and their tolerances.

NOTE Further information is provided in annex A.

**3.2 Limitations and restrictions on reprocessing**

The manufacturer shall determine if processing in accordance with the provided instructions leads to a degree of degradation that will limit the useful life of the medical device. Where such degradation is established, the manufacturer shall provide an indication of the number of reprocessing cycles that can normally be tolerated, or some other indication of the end of the medical device's ability to safely fulfil its intended use.

**3.3 Preparation at the point of use prior to processing**

Requirements for preparation at the point of use to ensure satisfactory reprocessing of the medical device, shall be specified, if applicable.

Where appropriate, at least the following information shall be included:

- the containers for transportation;
- a description of the support systems;
- the maximum period of time that may elapse between use and cleaning;
- a description of the pre-cleaning techniques critical to further processing;
- the requirements for transportation.

### 3.4 Preparation before cleaning

Requirements for the preparation of the medical device prior to cleaning shall be specified if applicable. Where appropriate, instructions for at least the following procedures shall be given:

- the requirements for capping/opening of ports;
- disassembly of the device;
- leak testing the device;
- soaking/brushing techniques required;
- ultrasonic treatment of the device.

If special tools are required for disassembly/re-assembly, these shall be specified in the instructions.

### 3.5 Cleaning

A validated method of manual cleaning shall be specified. At least one validated automated method using a washer-disinfector shall also be specified unless the medical device cannot withstand any such process, in which case a warning should be issued.

Where appropriate, at least the following information shall be included:

- a description of the accessories required for cleaning process;
- identification and concentration of chemicals required for cleaning;
- identification of water quality to be used for the process;
- limits and monitoring of chemical residues remaining on the device;
- limits on temperature, concentration of solution(s), exposure time to be used;
- the process temperature(s) to be used;
- the techniques to be used including rinsing.

**NOTE** Cleaning and Disinfecting Processing Equipment should be qualified and validated to ensure suitability for its intended purpose.

### 3.6 Disinfection

A validated non-automatic method of disinfection shall be specified. At least one validated automated method using a washer-disinfector shall also be specified unless the medical device cannot withstand any such process.

Where appropriate, at least the following information shall be included:

- a description of the accessories required for the disinfection process;
- the contact time of the disinfectant;
- identification and concentration of chemicals required for the disinfection process;
- identification of water quality required for the process;

- the limits and monitoring of chemical residues remaining on the device;
- the limits on temperature, concentration of solution(s), exposure time;
- the process temperature(s) to be used;
- the techniques to be used including rinsing.

NOTE 1 In certain circumstances disinfection may be carried out concurrently with cleaning of the medical device.

NOTE 2 Wherever practical a washer-disinfector using thermal disinfection is preferred.

NOTE 3 Certain clinical procedures lead to an enhanced probability of contamination with agents with high resistances against certain disinfectants (e.g. mycobacteria). This should be considered in the risk analysis and in the choice of recommended disinfectant.

### 3.7 Drying

Where drying is necessary, a validated method of drying shall be specified. Where appropriate at least the following information shall be included:

- the accessories required for the drying process;
- the maximum temperature and exposure time for the device;
- specifications of the drying agent to be used;
- the techniques to be used.

NOTE In certain circumstances drying may be achieved as part of an automated cleaning and disinfection process.

### 3.8 Inspection, maintenance and testing

When methods are required at any stage of processing to confirm the cleanliness or performance or both, of the medical device, these shall be stated. Where particular maintenance actions are required during processing to ensure the proper performance and safety of the medical device, these shall be stated. Where appropriate, these shall include details such as any part or component that requires routine replacement and/or calibration and where necessary, details for return to the manufacturer or other qualified organization.

Where appropriate at least the following information shall be given:

- the method to be used for adjustment/calibration of the device;
- a description of the lubrication to be used;
- the performance criteria for the device to ensure its safe use;
- the instructions for re-assembly of the device;
- the method to be used for the replacement of components;
- a description of special tools to be used to maintain the device;
- the requirements for visual inspection.

**NOTE** If instruction to return the medical device to the manufacturer or other qualified organization is specified, the manufacturer should give clear instructions as to the cleanliness and/or microbiological status required.

### 3.9 Packaging

If a specific method for packaging or containing the medical device during and after sterilization is required, it shall be stated and be compatible with the sterilization process and the medical device.

### 3.10 Sterilization

A validated method of sterilization shall be specified.

Where appropriate, at least the following information shall be given; this shall include set points and the upper and lower limits of critical process parameters that are capable of achieving sterility of the medical device:

- the accessories required for sterilization of the medical device;
- the identification and concentration of the sterilant required for the sterilization process;
- the identification of maximum values of contaminants in condensate from steam, used in moist heat, ethylene oxide and/or steam and formaldehyde sterilization;
- the humidity required for the sterilization process;
- the minimum holding or exposure time of sterilant;
- a description of post-sterilization techniques/activities;
- pressure required for the sterilization process;
- a description of the techniques to be used;
- the required temperature of the sterilant.

**NOTE** Wherever possible, moist heat sterilization is recommended.

### 3.11 Storage

Any specific limitations for the time or conditions of storage of the reprocessed medical device prior to use shall be stated.

## 4 Presentation of the information

**4.1** Where applicable, the information required by clause 3 shall accompany the medical device, e.g. in the instructions for use supplied with the medical device, or on the medical device label or packaging.

**NOTE** An example format for giving detailed information for a particular medical device is given in annex B.

**4.2** The information specified in clause 3 shall take into account the nature of the medical device, its intended use and the knowledge and training of the persons involved in the processing.

**NOTE** The information specified in clause 3 may make reference to:

- standards available;
- general processing information provided by the manufacturer;
- general processing information provided by the manufacturer of the equipment of materials involved in the specified process.

Such reference may be achieved by the use of symbols (ref: ISO 15223 and ISO 7000).

**4.3** The equipment or materials necessary in the specified processes shall be identified by its generic names or specification. Only in those cases where this does not provide sufficient information, trade names may be given in addition.

## **5 Validation of the reprocessing information provided**

The manufacturer shall validate that any process identified in the information provided is capable of reprocessing the medical device for its intended use.

**NOTE** Where the manufacturer supplies a number of different medical devices which share common features and attributes, the validation specified may be performed with respect to these medical devices as a group or family, provided that the manufacturer can demonstrate the commonality of the medical devices and that the tests and assessments address the “worst case” feature or attribute of the group or family.

## **6 Risk analysis**

In the risk analysis performed by the medical device manufacturer to determine the content and detail of the information to be provided, the medical device manufacturer shall take into account:

- the nature of the medical device;
- the intended use of the medical device;
- the likely training and knowledge of the processor;
- the equipment likely to be available to the processor.

**NOTE** See EN ISO 14937 as guidance.

## **Annex A** **(informative)**

### **Commonly utilized reprocessing methods**

#### **A.1 General**

The following example of a matrix is intended to assist the manufacturer of medical devices to identify methods of processing that may be considered for inclusion in the processing instructions provided.

The general acceptability of each of the methods for various categories of medical devices is indicated and may be used as a guide when considering the equipment, training and resources likely to be available to processors of the various medical device categories and so allow selection of processing methods most readily implemented by the processor.

Thorough cleaning prior to disinfection and sterilization is especially important for infection control.

This information also indicates what an experienced processor may assume to be appropriate reprocessing methods for certain medical device categories. As such it may be used as an input to the risk analysis required by this standard (clause 6) to determine the extent of warnings to avoid damaging or unsafe processing methods for a particular medical device.

The information in the following matrix is for guidance only and cannot be universally applied to all makes of medical devices within a category without reference to and compliance with the specific instructions provided.

IT IS STILL THE RESPONSIBILITY OF THE MANUFACTURER TO IDENTIFY AND VALIDATE SPECIFIC PROCEDURES FOR THE PARTICULAR MEDICAL DEVICE BEING CONSIDERED. LIKEWISE PROCESSORS SHOULD REFER TO, AND COMPLY WITH, THE SPECIFIC INSTRUCTIONS PROVIDED BY THE MANUFACTURER OF THE MEDICAL DEVICE, PROCESSING EQUIPMENT AND/OR PROCESSING CHEMICAL.

## A.2 Matrix for identifying methods of processing

Table A.1 — Examples of processes that might be applied

Process	Medical device									
	A	B	C	D	G	H	I	J		
	Reusable surgical instruments	Endoscopes (thermostable)	Endoscopes (thermolabile)	Instruments for use with endoscopes	Reusable containers	Elastic products <sup>1)</sup>	HF-cable and handpieces	Powertools		
Preparation at the point of use	+	+	+	+	-	+	+	+		
	+	+	+	+	+	+	+	+		
Decontamination	Selection according to the Cleaning and Disinfection Procedures									
	Preparation									
	Cleaning									
		Manual <sup>3)</sup>	+	+	+	+	+	+	+	
		Machine <sup>3)</sup>	+	+	+	+	+	+	+	
		Ultrasonic								
			-	+	N.A.	-	+	-		
		Alkaline detergent	+	+		+	+	+	-	
		Acidic detergent	-	-	-	-	+	+	-	
		Neutral detergent	+	+	+	+	+	+	+	
	Rinsing <sup>3)</sup>									
	Disinfection									
		Chemical	+	+	+	+	+	+	+	
	Thermal	+	+	+	+	+	+	+		
Drying										
	T <sub>max</sub>									

Table A.1 — Examples of processes that might be applied (concluded)

Maintenance																	
Sterilization		Moist heat (preferred method)	+	+	-	+	+	+	+	+	+	+	+	+	+	+	+
		Low temperature steam and fomaldehyde	+	+	+	+	+	+	+	+	-	+	+	+	+	+	-
		Ethylene oxide	+	+	+	+	+	+	+	+	+	-	+	+	+	+	-
		Dry heat	+	+							+						
		Liquid Steriliant	+	+	+	+	+	+	+	+	+	-	+	+	+	+	-
		Gas Plasma	+	+	+	+	+	+	+	+	+	+	+	+	+	+	+
		1) Rubber or latex (also combined), siliconelastomer, plastics, closed hollow bodies, thermolabile.															
		2) Applies only for aluminium alloys.															
		3) For final rinsing deionized water is recommended.															
		<b>Key</b>															
		+    might be suitable															
		-    not suitable															
		N.A.    not applicable															

NOTE The matrix provides typical results. Results may vary significantly depending on variations in device construction, device materials, device design and variations in reprocessing method.



Table A.2 — Table: blank table to define suitable reprocessing procedures

Process		Medical device							
		A	B	C	D	G	H	I	J
		Reusable surgical instruments	Endoscopes (thermostable)	Endoscopes (thermolabile)	Instruments for use with endoscopes	Reusable containers	Elastic products <sup>1)</sup>	HF-cable and handpieces	Power tools
Preparation at the point of use	Deposition	Wet Dry							
Decontamination	Preparation	Selection according to the Cleaning and Disinfection Procedures							
	Cleaning	Manual <sup>3)</sup>							
		Machine <sup>3)</sup>							
		Ultrasonic							
		Alkaline detergent							
		Acidic detergent							
		Neutral detergent							
	Rinsing <sup>3)</sup>								
	Disinfection	Chemical							
		Thermal							
	Drying	T <sub>max</sub>							

**Table A.2 — Table: blank table to define suitable reprocessing procedures (concluded)**

[illegible]

## **Annex B**

(informative)

### **An example of reprocessing instructions for reusable medical devices**

Processors may process medical devices from various device manufacturers, so for the sake of clarity manufacturers should adopt a consistent presentation of instructions for processing.

Processing instructions may be presented in accordance with Figure B.1 to aid manufacturers in achieving a consistent presentation.

The manufacturer should ensure that all required information is included, that it will be readily understood and the prominence of the various elements of the information is appropriate to their importance.

The following template Figure B.1 and example Figure B2 provide formats that may be used by manufacturers to achieve such consistency and may be applicable for the majority of medical devices.

**NOTE** This template represents one suggested format. There may be a number of different formats for the information that may be more appropriate.

**Manufacturer:** <Manufacturer name>      **Method :** <ref.>      **Symbol:** <sym>

**Device(s):** <list by catalogue number and device description, or generic type>

<b>WARNINGS</b>	<warnings re inappropriate chemicals, parameters, points of particular attention>
<b>Limitations on reprocessing</b>	<the number of reprocessing cycles permitted or other indications of end of life>

<b>INSTRUCTIONS</b>	
<b>Point of use:</b>	<instructions/cautions>
<b>Preparation for decontamination:</b>	<instructions/cautions>
<b>Cleaning: Automated</b>	<instructions/cautions. Include equipment/materials/parameters>
<b>Cleaning: Manual</b>	<instructions/cautions. Include equipment/materials/parameters>
<b>Disinfection:</b>	<instructions/cautions. Include equipment/materials/parameters>
<b>Drying</b>	<instructions/cautions, include equipment/materials/parameters>
<b>Maintenance, Inspection and Testing:</b>	<instructions/cautions. Include equipment/materials/parameters>
<b>Packaging:</b>	<instructions/cautions. Include materials/methods>
<b>Sterilization:</b>	<instructions/cautions. Include equipment/materials/parameters>
<b>Storage:</b>	<instructions/cautions>

<b>Additional Information:</b>	<Any other information considered helpful>
<b>Manufacturer contact:</b>	<Contact information for further information>

The instructions provided above have been validated by the manufacturer of the medical device as being CAPABLE of preparing a medical device for re-use. It remains the responsibility of the processor to ensure that the reprocessing as actually performed using equipment, materials and personnel in the reprocessing facility achieve the desired result. This normally requires validation and routine monitoring of the process.

Date issued: <date>

**Figure B.1 — Template: Processing instructions (reusable medical devices)**

Notes for use:

- 1) Instructions should be clear and concise.
- 2) Reference to materials and equipment should be generic where possible.
- 3) Instructions for disassembly/assembly, maintenance and inspection/test may be documented separately (these instructions are more likely to be specific to a particular medical device, whereas other instructions are more likely to apply to a group or family of medical devices).
- 4) All sections of the table require an entry. Phrases such as “no particular requirements”, “not applicable” etc. can be used where appropriate.
- 5) The symbol field is optional, it may be used to refer to the instructions from markings on the medical device or its packaging.

Manufacturer: I

Method:

Symbol:

Device(s): All reusable surgical instruments supplied by \_\_\_\_\_ Medical comprising fixed assemblies (no moving parts) and simple hinged assemblies, excluding those containing aluminium alloy.

NOTE Aluminium alloy may be recognized by bright coloured (red, blue, green, yellow) coatings on metallic components.

<b>WARNINGS</b>	<p>Aluminium based instruments are damaged by alkaline (pH&gt;7) detergents and solutions.</p> <p>Long narrow cannulations and blind holes require particular attention during cleaning.</p> <p>Do not exceed <b>yy</b>°C</p>
<b>Limitations on reprocessing</b>	Repeated processing has minimal effect on these instruments. End of life is normally determined by wear and damage due to use.

<b>INSTRUCTIONS</b>	
<b>Point of use:</b>	Remove excess soil with disposable cloth/paper wipe.
<b>Containment and transportation:</b>	<p>No particular requirements.</p> <p>It is recommended that instruments are reprocessed as soon as is reasonably practical following use.</p>
<b>Preparation for cleaning:</b>	<p>No particular requirements.</p> <p>Disassembly not required.</p>
<b>Cleaning: Automated</b>	<p>Equipment: Washer/disinfector, detergent (name)</p> <p>1 Load instruments such that hinges are open and cannulations and holes can drain.</p> <p>2 Run cycle, minimum <b>xx</b> minutes wash and <b>xx</b> minutes rinse.</p> <p>3 When unloading check cannulations, holes etc for complete removal of visible soil. If necessary repeat cycle or use manual cleaning.</p>
<b>Cleaning: Manual</b>	<p>Equipment: Detergent (example), brush, running water</p> <p>Method:</p> <p>1. Rinse excess soil from instrument.</p> <p>2. Using brush, apply detergent solution to all surfaces ensuring that hinged instruments are cleaned in both open and closed positions.</p> <p>NOTE Clean cannulations and holes using an appropriate brush ensuring that full depth of the feature is reached.</p> <p>3 Rinse under clean running water for <b>xx</b> minutes. Ensure that running water passes through cannulations, and that blind holes are repeatedly filled and emptied.</p>
<b>Disinfection:</b>	Disinfectant solution (name) may be used in accordance with label instructions.
	If automated cleaning is employed, a final rinse at <b>yy</b> °C for <b>xx</b> minutes may be used to effect thermal disinfection.

<b>Drying:</b>	When drying is achieved as part of a washer disinfector cycle do not exceed 120°C.
<b>Maintenance:</b>	Apply a small quantity of surgical grade lubrication oil to hinges. Discard blunt or damaged instruments.

<b>Inspection and Function Testing:</b>	<p>Hinged instruments: Check for smooth movement of hinge without excessive “play”. Locking (ratchet) mechanisms should be checked for action.</p> <p>All instruments: Visually inspect for damage and wear. Cutting edges should be free of nicks and present a continuous edge.</p> <p>Check instruments with long slender features (particularly rotating instruments) for distortion. Where instruments form part of a larger assembly, check assembly with mating components.</p>
<b>Packaging:</b>	<p>Singly: A standard packaging material may be used. Ensure that the pack is large enough to contain the instrument without stressing the seals.</p> <p>In sets: Instruments may be loaded into dedicated instrument trays, or general-purpose sterilization trays. Ensure that cutting edges are protected, and do not exceed <b>z</b> Kg per tray. Wrap the trays using appropriate method.</p>
<b>Sterilization:</b>	<p>Vacuum autoclave, minimum of <b>x</b> minutes at <b>y</b> °C.</p> <p>Do not exceed <b>y</b> °C.</p>
<b>Storage:</b>	No particular requirements.
<b>Additional Information:</b>	When sterilising multiple instruments in one autoclave cycle ensure that the sterilizer’s maximum load is not exceeded.
<b>Manufacturer contact:</b>	See brochure for telephone and address of local representative or telephone (44) 123 456 789.

The instructions provided above have been validated by the medical device manufacturer as being CAPABLE of preparing a medical device for re-use. It remains the responsibility of the processor to ensure that the processing as actually performed using equipment, materials and personnel in the processing facility achieve the desired result. This requires validation and routine monitoring of the process. Likewise any deviation by the processor from the instructions provided should be properly evaluated for effectiveness and potential adverse consequences.

**Figure B.2 — Example**

## Bibliography

- [1] ISO 11134:1994, *Sterilization of health care products — Requirements for validation and routine control — Industrial moist heat sterilization*
- [2] EN 285:1996, *Sterilization — Steam sterilizers — Large sterilizers*
- [3] ISO 11135:1994, *Medical Devices — Validation and routine control of ethylene oxide sterilization*
- [4] ISO 11607: 1997, *Packaging for terminally sterilized medical devices*
- [5] EN ISO 14937:2000, *Sterilization of health care products - General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices (ISO 14937:2000)*
- [6] ISO 15223:2000, *Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied*
- [7] ISO 7000:1989, *Graphical symbols for use on equipment – index and synopsis*
- [8] Medical Devices Agency: *Sterilization, Disinfection and Cleaning of Medical Equipment: Guidance on Decontamination from Microbiology Advisory Committee to the Department of Health, Medical Devices Agency; Part 1 Principles, Part 2 Protocols, Part 3 Procedures 1993-2000* London





